



Laerdal
helping save lives

DECLARATION OF CONFORMITY

Responsible Manufacturer: Laerdal Medical AS
Tanke Svilandsgate 30
P.O. Box 377
4002 Stavanger
Norway

Manufacturing site: Laerdal Texas
226 FM 116, P.O. Box 38
GATESVILLE, TEXAS 76528
United States of America

Product Name: SimNewB Standard, 220-000XX
SimNewB Advanced, 225-050XX

Product Options: Peripheral Kit Complete, 225-090XX
(w/Accessories) Peripheral Kit, 225-091XX

Note: XX = country code

to which this declaration relates is in conformity with the Essential Requirements of the Council Directive 1999/5/EC relating to Radio Equipment and Telecommunications Terminal Equipment.

All supporting documentation is retained by the manufacturer.

The conformity is based on the following standards:

EN 61000-6-1:2007, EN 61000-6-3:2007, EN 300 220-1 V.1.3.1:2000v, EN 300 220-3 V.1.1.1:2000, EN 301 489-1 V.1.4.1:2002, EN 301 489-3 V.1.4.1:2002, EN 60950, EN 55022, EN 55022, EN 61000-3-2, EN 61000-4-2, EN 61000-4-3, EN 61000-4-4, EN 61000-4-5, EN 61000-4-6 and, EN 61000-4-11 & EN 61010-1.

Stavanger, September 17 2008

Terje Bondhus
Regulatory Affairs Specialist

