

Declaration of Conformity

We,

POLYMED (XIAMEN) PLASTIC INDUSTRIAL CO., LTD.

UNIT B, 1~5 F, WAREHOUSE & PROCESS COMPLEX BUILDING XIANGYU F. T. Z.,
XIAMEN, CHINA

FAX : 86-592-5624688 TEL : 86-592-6024339

declare under our sole responsibility that the products listed below:

Item	RESPIRATORY CARE MEDICAL PRODUCTS		
Product	Resuscitator Unit		
Model	RUGH01/RUGH02		
Type/Group	Refer to Below Codes Listing		
Item No.	Name	Polymed Codes No.	Laerdal Codes No.
1	Laerdal Infant Disposable Resuscitator	1.RES.LAE.845031	845031
2	Laerdal Child Disposable Resuscitator	1.RES.LAE.845021	845021
3	Laerdal Adult Disposable Resuscitator	1.RES.LAE.845011	845011
4	Laerdal Adult Disposable Resuscitator w/ Mask #4	1.RES.LAE.845041	845041
5	Laerdal Disposable Mask #3 w/inflation port	1.CUS.LAE.845050	845050
6	Laerdal Disposable Mask #4 w/inflation port	1.CUS.LAE.845060	845060
7	Laerdal Disposable Mask #5 w/inflation port	1.CUS.LAE.845070	845070
8	Laerdal Disposable PEEP Valve (5-20cmH2O)	1.PEV.LAE.845040	845040
9	Laerdal Disposable Expiration Diverter	1.DIV.LAE.845080	845080

to which this declaration relates is in conformity with the following standards or other normative documents :

STANDARD	TEST REPORT/CERTIFICATE NO.	ISSURED BY
ISO 10993-5	C-49000	ITS Cranleigh (UK) Ltd.
ISO 10993-10	C-49000	ITS Cranleigh (UK) Ltd.
EN 980:2003	-----	-----
EN 1041:1998	-----	-----
ISO 10651-4:2002	-----	-----
ASTM F920-93,	-----	-----
EN 1441:1997	-----	-----
ISO 14971:2000	-----	-----
ISO 5356-1:2004	-----	-----
CMDCAS ISO 13485:2003	CERT-0022822	QMI Management Systems Registration
ISO 13485:2003	2006-OSL-AQ-0536	DET NORSKE VERITAS AS
CE MDD	2002-OSL-MDD-0177	DET NORSKE VERITAS AS

following the provisions of Directive 93/42/EEC Annex I, Annex V, Annex VII.

Issue at,

POLYMED (XIAMEN) PLASTIC INDUSTRIAL CO., LTD.

UNIT B, 1~5 F, BLOCK G, WAREHOUSE & PROCESS COMPLEX BUILDING XIANGYU F. T. Z.,
XIAMEN, CHINA

Date: 2007/04/11



Leo Chien
General Manager